
Midmark M11 - v1.0.0
Total Cycles: 2012

Name: RITTER 1
Sterilizer ID: V2178198
06/08/2022 03:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.3 F 0.0
2:00 197.1 F 4.6
4:00 203.9 F 9.8
6:00 221.4 F 6.2
8:00 233.7 F 11.1
10:00 233.7 F 8.4
12:00 240.1 F 10.3
14:00 246.5 F 13.5
16:00 259.1 F 20.4
18:00 270.3 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 29.1
2:00 271.6 F 28.3
3:00 272.6 F 29.2
4:00 272.1 F 28.6

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 230.9 F 0.0
10:00 242.0 F 0.0
15:00 238.9 F 0.0
20:00 228.7 F 0.0
25:00 226.7 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 18:50
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:07
CYCLE COMPLETE

06/08/2022 04:18 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____