
Midmark M11 - v1.0.0
Total Cycles: 5125

Name: RITTER 1
Sterilizer ID: V2178198
11/22/2024 09:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 71.5 F 0.0
4:00 86.6 F 0.0
6:00 108.3 F 0.0
8:00 136.2 F 0.1
10:00 168.8 F 0.3
12:00 192.1 F 0.7
14:00 208.2 F 3.7
16:00 215.8 F 6.4
18:00 224.1 F 9.5
20:00 232.6 F 13.2
22:00 233.5 F 8.7
24:00 243.8 F 13.4
26:00 241.1 F 11.1
28:00 236.8 F 9.1
30:00 250.3 F 15.4
32:00 261.2 F 21.5
34:00 271.1 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.7 F 29.0
2:00 271.8 F 28.2
3:00 272.1 F 28.7
4:00 272.5 F 28.7

Min 271.5 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.8 F 0.0
5:00 190.7 F 0.0
10:00 187.6 F 0.0
15:00 189.8 F 0.0
20:00 188.2 F 0.0
25:00 189.7 F 0.0
30:00 192.8 F 0.0
Drying Complete

FILLING: 0:42
HEATING: 34:46
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:12:58
CYCLE COMPLETE

11/22/2024 10:43 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____