
Midmark M11 - v1.0.0
Total Cycles: 5278

Name: RITTER 1
Sterilizer ID: V2178198
01/27/2025 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.4 F 0.0
2:00 73.2 F 0.0
4:00 85.0 F 0.1
6:00 126.1 F 2.1
8:00 151.5 F 4.9
10:00 175.5 F 8.2
12:00 190.7 F 11.7
14:00 202.4 F 2.8
16:00 214.4 F 6.4
18:00 226.0 F 10.2
20:00 230.3 F 8.4
22:00 238.1 F 10.6
24:00 234.6 F 7.9
26:00 247.9 F 14.4
28:00 259.5 F 20.7
30:00 270.2 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.1
2:00 271.6 F 28.5
3:00 272.8 F 29.1
4:00 271.8 F 28.4

Min 271.5 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 168.9 F 0.0
5:00 192.9 F 0.0
10:00 196.7 F 0.0
15:00 201.6 F 0.0
20:00 204.0 F 0.0
25:00 205.4 F 0.0
30:00 205.7 F 0.0
Drying Complete

FILLING: 0:43
HEATING: 30:53
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:09:04
CYCLE COMPLETE

01/27/2025 10:49 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____