
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000928
15:17 11/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:28
PRESSURIZING 3:31
STERILIZING 4:41
136.0 C 220kPa 4:41
136.4 C 232kPa 5:11
136.3 C 231kPa 5:41
136.3 C 232kPa 6:11
136.3 C 232kPa 6:41
136.3 C 232kPa 7:11
136.4 C 232kPa 7:41
136.4 C 231kPa 8:11
136.4 C 231kPa 8:11
VENTING 8:11
AIR DRYING 9:05
CYCLE COMPLETE 9:09

Digital Signature #
67C07238FC61365F
