
Midmark M11 - v1.0.0
Total Cycles: 2132

Name: RITTER 1
Sterilizer ID: V2178198
07/14/2022 03:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.6 F 0.0
2:00 132.8 F 0.0
4:00 164.5 F 2.4
6:00 184.1 F 6.3
8:00 208.1 F 11.0
10:00 218.7 F 4.5
12:00 230.1 F 8.6
14:00 238.0 F 13.1
16:00 237.9 F 10.0
18:00 237.5 F 8.9
20:00 247.8 F 14.0
22:00 258.6 F 20.1
24:00 269.2 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.2
1:00 272.4 F 29.0
2:00 271.6 F 28.1
3:00 272.4 F 29.0
4:00 271.9 F 28.3

Min 271.3 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.1 F 0.0
5:00 201.2 F 0.0
10:00 201.3 F 0.0
15:00 218.9 F 0.0
20:00 213.6 F 0.0
25:00 214.9 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 24:57
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:03:27
CYCLE COMPLETE

07/14/2022 04:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____