
Midmark M11 - v1.0.0
Total Cycles: 4974

Name: RITTER 1
Sterilizer ID: V2178198
10/07/2024 04:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.6 F 0.0
2:00 173.7 F 3.6
4:00 210.8 F 8.6
6:00 224.8 F 7.4
8:00 232.0 F 9.3
10:00 234.8 F 8.7
12:00 244.8 F 12.8
14:00 242.6 F 11.4
16:00 256.2 F 18.4
18:00 268.0 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 28.7
2:00 271.4 F 28.4
3:00 272.5 F 28.6
4:00 271.3 F 28.3

Min 271.2 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.5 F 0.0
5:00 232.1 F 0.0
10:00 224.1 F 0.0
15:00 236.0 F 0.0
20:00 230.2 F 0.0
25:00 226.4 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 19:14
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:58:00
CYCLE COMPLETE

10/07/2024 05:30 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____