
Midmark M11 - v1.0.0
Total Cycles: 1333

Name: RITTER 1
Sterilizer ID: V2178198
11/03/2021 06:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.4 F 0.0
2:00 74.0 F 0.0
4:00 78.7 F 0.0
6:00 100.5 F 0.1
8:00 135.9 F 0.1
10:00 167.2 F 0.5
12:00 199.5 F 3.2
14:00 200.8 F 5.8
16:00 211.5 F 8.8
18:00 222.2 F 12.2
20:00 227.8 F 6.7
22:00 239.4 F 11.3
24:00 234.5 F 8.3
26:00 244.0 F 11.6
28:00 245.8 F 12.9
30:00 257.4 F 19.0
32:00 267.7 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.0
1:00 271.9 F 28.5
2:00 272.5 F 28.4
3:00 271.4 F 27.9
4:00 272.7 F 28.8

Min 271.4 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.6 F 0.0
5:00 224.4 F 0.0
10:00 197.7 F 0.0
15:00 194.5 F 0.0
20:00 221.3 F 0.0
25:00 226.4 F 0.0
30:00 231.4 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 33:19
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:11:29
CYCLE COMPLETE

11/03/2021 07:58 AM
ACCEPTED: X
REJECTED:
Operator: U01