
Midmark M11 - v1.0.0
Total Cycles: 5289

Name: RITTER 1
Sterilizer ID: V2178198
01/28/2025 06:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.2 F 0.0
2:00 99.9 F 0.0
4:00 126.7 F 2.6
6:00 155.6 F 5.4
8:00 178.7 F 8.7
10:00 197.1 F 12.1
12:00 212.0 F 4.1
14:00 218.9 F 8.3
16:00 231.3 F 12.4
18:00 232.3 F 8.1
20:00 243.0 F 13.2
22:00 241.1 F 11.1
24:00 252.7 F 16.9
26:00 262.5 F 22.8
28:00 272.1 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.4 F 29.1
2:00 272.4 F 29.5
3:00 271.8 F 29.0
4:00 272.3 F 29.2

Min 271.7 F 28.9
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.9 F 0.0
5:00 201.6 F 0.0
10:00 205.7 F 0.0
15:00 197.5 F 0.0
20:00 204.4 F 0.0
25:00 204.4 F 0.0
30:00 205.3 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 28:28
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:07:11
CYCLE COMPLETE

01/28/2025 08:05 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____