

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4389

Name: RITTER 1  
Sterilizer ID: V2178198  
03/21/2024 08:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.8 F 0.0  
2:00 90.9 F 0.2  
4:00 100.3 F 0.6  
6:00 108.7 F 2.9  
8:00 138.0 F 5.8  
10:00 162.9 F 9.2  
12:00 185.7 F 12.9  
14:00 209.9 F 3.7  
16:00 219.6 F 6.8  
18:00 227.5 F 10.4  
20:00 231.8 F 8.6  
22:00 238.5 F 10.3  
24:00 236.2 F 8.5  
26:00 246.8 F 13.5  
28:00 257.9 F 19.5  
30:00 268.1 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.6 F 28.8  
2:00 271.4 F 28.0  
3:00 272.5 F 28.9  
4:00 272.0 F 28.4  
  
Min 271.2 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.5 F 0.0  
5:00 231.7 F 0.0  
10:00 225.7 F 0.0  
15:00 222.2 F 0.0  
20:00 225.3 F 0.0  
25:00 230.2 F 0.0  
30:00 226.4 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 31:23  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:09:39  
CYCLE COMPLETE

03/21/2024 09:40 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_