
Midmark M11 - v1.0.5
Total Cycles: 4325

Name: RITTER2
Sterilizer ID: V2313228
01/31/2025 03:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 113.9 F 0.0
2:00 113.0 F 0.9
4:00 131.1 F 3.2
6:00 162.0 F 6.6
8:00 183.3 F 10.0
10:00 202.2 F 11.0
12:00 222.3 F 6.0
14:00 232.3 F 9.9
16:00 237.0 F 10.0
18:00 238.7 F 10.6
20:00 236.7 F 8.8
22:00 247.5 F 14.3
24:00 258.1 F 20.1
26:00 267.8 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.3
1:00 272.8 F 29.6
2:00 272.5 F 29.8
3:00 271.8 F 29.8
4:00 271.7 F 29.2

Min 271.5 F 29.2
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.2 F 0.0
5:00 236.1 F 0.0
10:00 220.0 F 0.0
15:00 209.5 F 0.0
20:00 193.1 F 0.0
25:00 193.3 F 0.0
30:00 181.0 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 27:59
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:06:51
CYCLE COMPLETE

01/31/2025 05:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____