
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002047
12:56 16/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:06
STERILIZING 3:14
136.4 C 225kPa 3:14
136.4 C 226kPa 3:44
136.4 C 225kPa 4:14
136.4 C 225kPa 4:44
136.4 C 225kPa 5:14
136.3 C 225kPa 5:44
136.2 C 223kPa 6:14
136.4 C 223kPa 6:44
136.4 C 223kPa 6:44
VENTING 6:44
AIR DRYING 7:35
CYCLE COMPLETE 8:11

Digital Signature #
53F0B8085820A38F
