

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002629  
8:41 28/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:24  
PRESSURIZING 2:33  
STERILIZING 3:24  
136.0 C 220kPa 3:24  
136.3 C 224kPa 3:54  
136.4 C 225kPa 4:24  
136.4 C 222kPa 4:54  
136.2 C 223kPa 5:24  
136.3 C 224kPa 5:54  
136.4 C 223kPa 6:24  
136.1 C 223kPa 6:54  
136.1 C 223kPa 6:54  
VENTING 6:54  
AIR DRYING 7:44  
CYCLE COMPLETE 9:26  
-----

Digital Signature #  
33E6C692CC977926  
-----