
Midmark M11 - v1.0.5
Total Cycles: 1784

Name: RITTER2
Sterilizer ID: V2313228
11/29/2022 11:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.8 F 0.0
2:00 163.6 F 0.2
4:00 186.9 F 0.9
6:00 204.9 F 5.4
8:00 222.0 F 10.6
10:00 228.1 F 6.9
12:00 241.8 F 13.3
14:00 242.0 F 12.1
16:00 242.4 F 11.7
18:00 256.5 F 19.1
20:00 268.6 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.5 F 29.3
2:00 271.6 F 28.8
3:00 272.7 F 29.5
4:00 272.0 F 28.8

Min 271.5 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.3 F 0.0
5:00 217.4 F 0.0
10:00 229.9 F 0.0
15:00 224.9 F 0.0
20:00 224.5 F 0.0
25:00 225.6 F 0.0
30:00 225.6 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 21:08
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:15
CYCLE COMPLETE

11/29/2022 12:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____