
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000427
13:38 27/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:34
STERILIZING 3:36
136.5 C 226kPa 3:36
136.4 C 233kPa 4:06
136.4 C 232kPa 4:36
136.2 C 232kPa 5:06
136.3 C 232kPa 5:36
135.9 C 231kPa 6:06
136.3 C 233kPa 6:36
136.3 C 233kPa 7:06
136.3 C 233kPa 7:06
VENTING 7:06
AIR DRYING 8:00
CYCLE COMPLETE 8:32

Digital Signature #
C52AAA0B0C6AFA42
