
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000425
12:46 27/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:36
STERILIZING 3:37
136.4 C 220kPa 3:37
136.3 C 232kPa 4:07
136.4 C 233kPa 4:37
136.2 C 231kPa 5:07
136.4 C 233kPa 5:37
136.4 C 231kPa 6:07
136.3 C 233kPa 6:37
136.4 C 232kPa 7:07
136.4 C 232kPa 7:07
VENTING 7:07
AIR DRYING 7:59
CYCLE COMPLETE 8:09

Digital Signature #
97932E2072C79C2B
