
Midmark M11 - v1.0.0
Total Cycles: 1034

Name: RITTER 1
Sterilizer ID: V2178198
07/21/2021 06:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.8 F 0.0
2:00 72.9 F 0.0
4:00 76.6 F 0.1
6:00 96.5 F 0.1
8:00 134.3 F 0.1
10:00 171.2 F 0.2
12:00 195.1 F 0.3
14:00 216.9 F 3.2
16:00 224.4 F 5.8
18:00 231.7 F 8.6
20:00 239.6 F 12.1
22:00 234.2 F 8.4
24:00 245.0 F 13.5
26:00 242.2 F 11.4
28:00 239.0 F 9.8
30:00 251.8 F 16.0
32:00 262.6 F 22.1
34:00 272.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.6 F 28.6
2:00 271.5 F 28.1
3:00 272.6 F 28.6
4:00 271.3 F 28.0

Min 271.3 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.4 F 0.0
5:00 213.1 F 0.0
10:00 216.3 F 0.0
15:00 209.1 F 0.0
20:00 227.7 F 0.0
25:00 238.4 F 0.0
30:00 204.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 34:27
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:12:36
CYCLE COMPLETE

07/21/2021 07:51 AM
ACCEPTED: X
REJECTED:
Operator: U01