
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 008206
14:29 09/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:42
STERILIZING 3:48
136.4 C 224kPa 3:48
135.9 C 221kPa 4:18
136.4 C 224kPa 4:48
136.2 C 224kPa 5:18
136.4 C 223kPa 5:48
136.4 C 224kPa 6:18
136.1 C 222kPa 6:48
136.2 C 223kPa 7:18
136.2 C 223kPa 7:18
VENTING 7:18
AIR DRYING 8:07
CYCLE COMPLETE 19:17

Digital Signature #
52E4BB4220C784A7
