
Midmark M11 - v1.0.5
Total Cycles: 2663

Name: RITTER2
Sterilizer ID: V2313228
07/18/2023 12:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.9 F 0.0
2:00 192.7 F 3.7
4:00 197.4 F 8.6
6:00 214.0 F 9.5
8:00 227.4 F 9.5
10:00 233.5 F 9.0
12:00 242.7 F 13.2
14:00 243.0 F 12.4
16:00 256.4 F 19.4
18:00 268.0 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 271.6 F 29.5
2:00 272.6 F 29.8
3:00 272.0 F 29.3
4:00 271.3 F 29.0

Min 271.2 F 28.9
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 208.4 F 0.0
10:00 225.0 F 0.0
15:00 235.7 F 0.0
20:00 226.2 F 0.0
25:00 221.2 F 0.0
30:00 217.0 F 0.0

Drying Complete

FILLING: 1:48
HEATING: 19:11
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:13
CYCLE COMPLETE

07/18/2023 01:35 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____