

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000563  
15:31 04/08/2021  
WRAPPED  
135 C for 10 min.  
CYCLE START 0:00  
CONDITIONING 2:48  
PRESSURIZING 5:44  
STERILIZING 6:48  
136.4 C 233kPa 6:48  
136.4 C 232kPa 7:18  
136.2 C 230kPa 7:48  
136.4 C 229kPa 8:18  
136.3 C 232kPa 8:48  
136.0 C 229kPa 9:18  
136.4 C 233kPa 9:48  
136.2 C 231kPa 10:18  
136.4 C 231kPa 10:48  
136.4 C 230kPa 11:18  
136.3 C 232kPa 11:48  
136.4 C 232kPa 12:18  
136.4 C 232kPa 12:48  
136.2 C 230kPa 13:18  
136.2 C 230kPa 13:48  
136.3 C 232kPa 14:18  
136.4 C 233kPa 14:48  
136.4 C 233kPa 15:18  
136.1 C 231kPa 15:48  
136.4 C 233kPa 16:18  
136.4 C 230kPa 16:48  
136.4 C 230kPa 16:48  
VENTING 16:48  
AIR DRYING 17:43  
CYCLE COMPLETE 32:43  
-----

Digital Signature #  
7BBD658E0C272948  
-----