

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001740  
11:34 17/04/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:06  
PRESSURIZING 2:11  
STERILIZING 3:13  
136.0 C 222kPa 3:13  
135.9 C 221kPa 3:43  
136.2 C 222kPa 4:13  
136.2 C 222kPa 4:43  
136.1 C 222kPa 5:13  
136.1 C 222kPa 5:43  
136.3 C 224kPa 6:13  
136.2 C 222kPa 6:43  
136.2 C 222kPa 6:43  
VENTING 6:43  
AIR DRYING 7:34  
CYCLE COMPLETE 9:17  
-----

Digital Signature #  
42EE869654024972  
-----