
Midmark M11 - v1.0.0
Total Cycles: 4908

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2024 12:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.9 F 0.0
2:00 182.4 F 0.3
4:00 198.9 F 2.1
6:00 214.1 F 6.6
8:00 228.2 F 11.3
10:00 229.8 F 6.7
12:00 240.9 F 12.2
14:00 239.4 F 10.0
16:00 237.5 F 8.6
18:00 250.5 F 15.1
20:00 261.9 F 21.6
22:00 272.1 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.3 F 28.7
2:00 272.6 F 28.7
3:00 272.2 F 28.3
4:00 271.5 F 28.0

Min 271.2 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.0 F 0.0
5:00 222.8 F 0.0
10:00 196.0 F 0.0
15:00 186.4 F 0.0
20:00 182.1 F 0.0
25:00 180.8 F 0.0
30:00 180.4 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 22:38
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:01:19
CYCLE COMPLETE

09/13/2024 01:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____