
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001776
15:42 15/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:45
STERILIZING 3:50
136.4 C 225kPa 3:50
136.2 C 223kPa 4:20
136.4 C 224kPa 4:50
136.2 C 222kPa 5:20
136.3 C 223kPa 5:50
136.2 C 223kPa 6:20
136.4 C 225kPa 6:50
136.3 C 224kPa 7:20
136.3 C 224kPa 7:20
VENTING 7:20
AIR DRYING 8:10
CYCLE COMPLETE 29:06

Digital Signature #
390589BF50C4A6E1
