
Midmark M11 - v1.0.5
Total Cycles: 2031

Name: RITTER2
Sterilizer ID: V2313228
02/09/2023 11:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.1 F 0.0
2:00 189.6 F 3.6
4:00 203.8 F 8.9
6:00 219.0 F 8.4
8:00 232.3 F 10.8
10:00 234.0 F 8.9
12:00 238.1 F 9.7
14:00 249.5 F 15.4
16:00 262.4 F 22.8
18:00 272.3 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.2 F 29.4
2:00 271.3 F 29.0
3:00 272.3 F 29.6
4:00 271.5 F 28.9

Min 271.2 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.9 F 0.0
5:00 217.9 F 0.0
10:00 230.2 F 0.0
15:00 233.4 F 0.0
20:00 222.6 F 0.0
25:00 216.5 F 0.0
30:00 218.3 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 18:11
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:12
CYCLE COMPLETE

02/09/2023 12:56 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____