
Midmark M11 - v1.0.5
Total Cycles: 3018

Name: RITTER2
Sterilizer ID: V2313228
10/17/2023 11:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.7 F 0.0
2:00 164.2 F 3.2
4:00 181.1 F 7.3
6:00 203.5 F 12.5
8:00 216.7 F 7.1
10:00 230.9 F 12.5
12:00 236.9 F 9.9
14:00 236.2 F 8.7
16:00 249.6 F 15.4
18:00 262.0 F 22.4
20:00 272.9 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.7 F 29.6
2:00 272.4 F 29.3
3:00 272.8 F 29.9
4:00 271.8 F 29.2

Min 271.0 F 28.7
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.7 F 0.0
5:00 201.3 F 0.0
10:00 206.8 F 0.0
15:00 205.7 F 0.0
20:00 202.4 F 0.0
25:00 199.2 F 0.0
30:00 195.6 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 20:22
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:49
CYCLE COMPLETE

10/17/2023 12:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____