

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001575  
11:35 11/11/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:13  
PRESSURIZING 6:59  
STERILIZING 8:40  
136.4 C 223kPa 8:40  
136.5 C 223kPa 9:10  
135.7 C 220kPa 9:40  
136.1 C 221kPa 10:10  
136.3 C 221kPa 10:40  
136.4 C 222kPa 11:10  
136.4 C 220kPa 11:40  
136.3 C 221kPa 12:10  
VENTING 12:10  
AIR DRYING 13:16  
CYCLE COMPLETE 43:16  
-----

Digital Signature #  
AADA76FC93409DEC  
-----