
Midmark M11 - v1.0.5
Total Cycles: 1450

Name: RITTER2
Sterilizer ID: V2313228
09/02/2022 01:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 213.1 F 0.0
2:00 209.9 F 4.3
4:00 220.8 F 11.0
6:00 230.1 F 8.5
8:00 237.2 F 9.9
10:00 242.0 F 11.4
12:00 247.9 F 14.5
14:00 261.1 F 21.9
16:00 272.5 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.6 F 29.4
2:00 271.5 F 28.8
3:00 272.7 F 29.5
4:00 272.0 F 28.9

Min 271.3 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 230.2 F 0.0
10:00 242.4 F 0.0
15:00 229.1 F 0.0
20:00 222.5 F 0.0
Dry Cycle Aborted

FILLING: 2:45
HEATING: 16:26
STERILIZING: 4:00
VENTING: 3:04
DRYING: 21:39
TOTAL CYCLE: 00:47:55

09/02/2022 02:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____