
Midmark M11 - v1.0.0
Total Cycles: 2281

Name: RITTER 1
Sterilizer ID: V2178198
09/02/2022 09:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.7 F 0.0
2:00 190.1 F 3.5
4:00 196.6 F 8.0
6:00 222.3 F 12.9
8:00 226.5 F 7.3
10:00 236.6 F 12.4
12:00 235.7 F 9.3
14:00 239.5 F 9.5
16:00 247.0 F 13.6
18:00 258.7 F 19.9
20:00 269.2 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.3 F 29.1
2:00 272.3 F 28.7
3:00 271.5 F 28.2
4:00 271.6 F 28.6

Min 271.2 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 230.8 F 0.0
10:00 240.5 F 0.0
15:00 234.6 F 0.0
20:00 228.9 F 0.0
25:00 228.9 F 0.0
30:00 236.5 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 21:11
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:22
CYCLE COMPLETE

09/02/2022 10:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____