
Midmark M11 - v1.0.0
Total Cycles: 4845

Name: RITTER 1
Sterilizer ID: V2178198
08/22/2024 03:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.8 F 0.0
2:00 161.0 F 0.7
4:00 176.4 F 2.6
6:00 202.5 F 7.5
8:00 218.8 F 12.9
10:00 226.8 F 7.8
12:00 238.5 F 13.4
14:00 239.5 F 10.9
16:00 237.2 F 9.3
18:00 252.3 F 16.5
20:00 264.6 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.3 F 28.8
2:00 271.6 F 28.6
3:00 272.2 F 28.7
4:00 271.5 F 28.5

Min 271.3 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 210.9 F 0.0
10:00 212.6 F 0.0
15:00 219.8 F 0.0
20:00 231.6 F 0.0
25:00 230.8 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 21:49
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:00:09
CYCLE COMPLETE

08/22/2024 04:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____