
Midmark M11 - v1.0.5
Total Cycles: 1805

Name: RITTER2
Sterilizer ID: V2313228
12/05/2022 05:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.2 F 0.0
2:00 161.1 F 3.3
4:00 175.4 F 7.1
6:00 197.7 F 12.1
8:00 219.9 F 6.3
10:00 230.3 F 11.4
12:00 232.3 F 8.7
14:00 239.2 F 10.3
16:00 246.6 F 14.1
18:00 259.1 F 21.0
20:00 270.8 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.4 F 29.8
2:00 271.2 F 29.0
3:00 272.3 F 29.9
4:00 272.0 F 29.4

Min 271.1 F 29.0
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.9 F 0.0
5:00 234.2 F 0.0
10:00 247.7 F 0.0
15:00 222.5 F 0.0
20:00 216.4 F 0.0
25:00 217.1 F 0.0
30:00 218.6 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 20:47
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:20
CYCLE COMPLETE

12/05/2022 06:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____