
Midmark M11 - v1.0.5
Total Cycles: 2522

Name: RITTER2
Sterilizer ID: V2313228
06/15/2023 12:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.7 F 0.0
2:00 183.0 F 4.0
4:00 201.8 F 9.4
6:00 220.2 F 7.9
8:00 232.0 F 10.9
10:00 232.9 F 8.5
12:00 239.4 F 10.4
14:00 247.2 F 14.5
16:00 259.9 F 21.5
18:00 271.1 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.7 F 29.9
2:00 271.9 F 29.2
3:00 271.7 F 29.5
4:00 272.8 F 29.9

Min 271.4 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 209.5 F 0.0
10:00 219.8 F 0.0
15:00 223.4 F 0.0
20:00 223.1 F 0.0
25:00 221.3 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 18:44
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:57:56
CYCLE COMPLETE

06/15/2023 01:05 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____