
Midmark M11 - v1.0.0
Total Cycles: 3057

Name: RITTER 1
Sterilizer ID: V2178198
03/30/2023 09:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.4 F 0.0
2:00 193.6 F 3.9
4:00 191.6 F 8.6
6:00 211.8 F 10.1
8:00 225.3 F 8.7
10:00 235.0 F 10.5
12:00 240.5 F 11.5
14:00 239.4 F 10.0
16:00 253.2 F 16.9
18:00 265.1 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.5 F 29.1
2:00 272.0 F 28.5
3:00 271.5 F 28.5
4:00 272.6 F 29.2

Min 271.3 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 218.6 F 0.0
10:00 233.3 F 0.0
15:00 234.8 F 0.0
20:00 228.0 F 0.0
25:00 223.2 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 19:50
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:32
CYCLE COMPLETE

03/30/2023 10:52 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____