
Midmark M11 - v1.0.5
Total Cycles: 2648

Name: RITTER2
Sterilizer ID: V2313228
07/14/2023 01:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.5 F 0.0
2:00 117.9 F 0.0
4:00 138.5 F 0.7
6:00 170.5 F 4.5
8:00 190.1 F 8.5
10:00 208.3 F 12.5
12:00 218.4 F 6.2
14:00 229.8 F 10.5
16:00 233.0 F 8.6
18:00 241.4 F 12.3
20:00 240.4 F 10.9
22:00 252.6 F 17.1
24:00 263.4 F 23.5
26:00 272.5 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.8
2:00 271.8 F 29.1
3:00 271.4 F 29.2
4:00 272.6 F 29.8

Min 271.2 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.1 F 0.0
5:00 208.1 F 0.0
10:00 209.8 F 0.0
15:00 209.4 F 0.0
20:00 207.8 F 0.0
25:00 207.4 F 0.0
30:00 206.6 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 26:15
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:04:31
CYCLE COMPLETE

07/14/2023 02:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____