
Midmark M11 - v1.0.0
Total Cycles: 2111

Name: RITTER 1
Sterilizer ID: V2178198
07/11/2022 09:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.4 F 0.0
2:00 204.2 F 2.5
4:00 207.0 F 6.4
6:00 217.4 F 11.2
8:00 224.7 F 5.7
10:00 233.9 F 11.0
12:00 234.3 F 7.8
14:00 243.5 F 11.9
16:00 243.6 F 12.3
18:00 255.7 F 18.7
20:00 266.6 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.6 F 29.3
2:00 271.5 F 28.5
3:00 272.3 F 29.4
4:00 272.1 F 28.9

Min 271.3 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.2 F 0.0
5:00 218.2 F 0.0
10:00 213.8 F 0.0
15:00 212.3 F 0.0
20:00 230.9 F 0.0
25:00 227.3 F 0.0
30:00 230.0 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 21:38
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:56
CYCLE COMPLETE

07/11/2022 10:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____