
Midmark M11 - v1.0.0
Total Cycles: 5309

Name: RITTER 1
Sterilizer ID: V2178198
02/03/2025 01:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 108.8 F 0.0
2:00 115.5 F 1.1
4:00 134.3 F 4.3
6:00 162.8 F 7.4
8:00 185.5 F 11.0
10:00 211.1 F 3.4
12:00 214.8 F 6.0
14:00 226.5 F 9.8
16:00 234.2 F 9.6
18:00 238.0 F 10.1
20:00 237.8 F 9.1
22:00 246.6 F 13.5
24:00 257.6 F 19.5
26:00 268.2 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.4
1:00 271.8 F 28.8
2:00 272.4 F 29.0
3:00 271.8 F 28.4
4:00 271.6 F 28.7

Min 271.2 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.5 F 0.0
5:00 224.4 F 0.0
10:00 211.7 F 0.0
15:00 221.1 F 0.0
20:00 239.2 F 0.0
25:00 235.9 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 27:16
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:05:54
CYCLE COMPLETE

02/03/2025 02:26 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____