

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2623

Name: RITTER 1  
Sterilizer ID: V2178198  
11/29/2022 05:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.4 F 0.0  
2:00 188.0 F 4.6  
4:00 199.2 F 9.6  
6:00 217.0 F 6.8  
8:00 228.0 F 9.9  
10:00 233.3 F 8.0  
12:00 242.5 F 13.0  
14:00 242.9 F 11.8  
16:00 256.3 F 18.9  
18:00 268.2 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.5 F 29.0  
2:00 271.3 F 28.3  
3:00 272.6 F 29.2  
4:00 272.1 F 28.7  
  
Min 271.3 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.1 F 0.0  
5:00 231.1 F 0.0  
10:00 238.5 F 0.0  
15:00 238.7 F 0.0  
20:00 229.9 F 0.0  
25:00 227.7 F 0.0  
30:00 229.2 F 0.0  
Drying Complete

FILLING: 1:41  
HEATING: 19:12  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:05  
CYCLE COMPLETE

11/29/2022 06:38 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_