
Midmark M11 - v1.0.5
Total Cycles: 2168

Name: RITTER2
Sterilizer ID: V2313228
03/22/2023 12:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.5 F 0.0
2:00 185.0 F 3.7
4:00 203.3 F 9.0
6:00 219.7 F 8.2
8:00 228.7 F 11.3
10:00 235.1 F 9.5
12:00 236.5 F 9.0
14:00 250.0 F 15.9
16:00 262.8 F 23.1
18:00 271.5 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 271.6 F 29.6
2:00 271.8 F 29.8
3:00 271.8 F 29.8
4:00 271.8 F 29.8

Min 271.1 F 29.2
Max 272.3 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.8 F 0.0
5:00 238.1 F 0.0
10:00 232.0 F 0.0
15:00 230.9 F 0.0
20:00 228.6 F 0.0
25:00 228.0 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 18:02
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:57:34
CYCLE COMPLETE

03/22/2023 01:25 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____