

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000235  
14:40 28/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:07  
PRESSURIZING 3:23  
STERILIZING 5:07  
136.4 C 233kPa 5:07  
136.4 C 233kPa 5:37  
136.4 C 233kPa 6:07  
136.4 C 233kPa 6:37  
136.3 C 233kPa 7:07  
136.1 C 232kPa 7:37  
136.4 C 232kPa 8:07  
136.3 C 233kPa 8:37  
136.3 C 233kPa 8:37  
VENTING 8:37  
CYCLE ABORTED 9:45  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 9:45  
-----

Digital Signature #  
5B3367C468CAB9A5  
-----