
Midmark M11 - v1.0.5
Total Cycles: 2106

Name: RITTER2
Sterilizer ID: V2313228
03/01/2023 08:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.5 F 0.0
2:00 81.1 F 0.0
4:00 117.6 F 0.0
6:00 141.4 F 0.2
8:00 168.7 F 0.4
10:00 175.6 F 3.3
12:00 195.2 F 6.7
14:00 217.8 F 10.5
16:00 229.1 F 7.3
18:00 231.2 F 9.2
20:00 239.9 F 11.6
22:00 238.9 F 11.0
24:00 236.7 F 9.1
26:00 249.5 F 15.4
28:00 260.5 F 21.6
30:00 270.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.6 F 29.5
2:00 271.3 F 29.0
3:00 272.5 F 29.6
4:00 271.5 F 28.9

Min 271.3 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 208.0 F 0.0
10:00 226.3 F 0.0
15:00 229.9 F 0.0
20:00 228.3 F 0.0
25:00 229.0 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:49
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:08:51
CYCLE COMPLETE

03/01/2023 09:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____