
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000794
10:51 30/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:39
STERILIZING 3:44
136.4 C 233kPa 3:44
136.4 C 233kPa 4:14
136.1 C 231kPa 4:44
136.2 C 232kPa 5:14
136.2 C 232kPa 5:44
136.4 C 233kPa 6:14
136.4 C 233kPa 6:44
136.4 C 233kPa 7:14
136.4 C 233kPa 7:14
VENTING 7:14
AIR DRYING 8:06
CYCLE COMPLETE 10:19

Digital Signature #
8F9D0CD6A88CE3DB
