
Midmark M11 - v1.0.5
Total Cycles: 4486

Name: RITTER2
Sterilizer ID: V2313228
04/04/2025 08:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 75.7 F 0.1
4:00 109.1 F 0.6
6:00 104.6 F 2.6
8:00 142.1 F 5.4
10:00 162.2 F 8.9
12:00 182.0 F 12.8
14:00 201.7 F 3.7
16:00 214.3 F 7.1
18:00 225.6 F 10.9
20:00 227.8 F 7.1
22:00 238.8 F 11.6
24:00 237.0 F 9.4
26:00 249.9 F 15.7
28:00 260.8 F 21.9
30:00 270.8 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.9 F 29.6
2:00 271.9 F 29.2
3:00 272.8 F 29.5
4:00 271.9 F 29.2

Min 271.6 F 28.9
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.9 F 0.0
5:00 180.0 F 0.0
10:00 194.0 F 0.0
15:00 202.8 F 0.0
20:00 203.4 F 0.0
25:00 205.2 F 0.0
30:00 203.3 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:47
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:09:05
CYCLE COMPLETE

04/04/2025 09:59 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____