
Midmark M11 - v1.0.0
Total Cycles: 698

Name: MIDMARK
Sterilizer ID: V2178198
02/24/2021 12:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.0 F 0.0
2:00 163.7 F 0.8
4:00 166.9 F 3.6
6:00 181.4 F 7.8
8:00 201.9 F 12.9
10:00 223.5 F 7.1
12:00 237.0 F 12.4
14:00 237.4 F 9.3
16:00 236.4 F 8.8
18:00 247.6 F 13.9
20:00 259.9 F 20.7
22:00 271.1 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 271.7 F 28.6
2:00 271.9 F 28.6
3:00 272.2 F 28.9
4:00 272.7 F 29.2

Min 271.4 F 28.6
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.1 F 0.0
5:00 234.8 F 0.0
10:00 239.1 F 0.0
15:00 226.4 F 0.0
20:00 219.2 F 0.0
25:00 219.5 F 0.0
30:00 218.0 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 23:21
STERILIZING: 4:00
VENTING: 2:56
DRYING: 30:00
TOTAL CYCLE: 01:01:10
CYCLE COMPLETE

02/24/2021 01:57 PM
ACCEPTED: X
REJECTED:
Operator: U01