
Midmark M11 - v1.0.0
Total Cycles: 5499

Name:
Sterilizer ID: V2178198
04/11/2025 11:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.2 F 0.0
2:00 174.9 F 3.7
4:00 189.5 F 8.0
6:00 209.0 F 12.8
8:00 221.4 F 6.6
10:00 231.0 F 11.3
12:00 233.2 F 7.8
14:00 245.1 F 13.3
16:00 243.2 F 11.7
18:00 254.5 F 17.5
20:00 264.5 F 23.5
22:00 271.6 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 28.8
2:00 271.1 F 28.1
3:00 272.5 F 29.2
4:00 272.1 F 28.6

Min 270.9 F 28.0
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 205.1 F 0.0
10:00 197.2 F 0.0
15:00 226.2 F 0.0
20:00 234.3 F 0.0
25:00 230.4 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 22:06
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:01:15
CYCLE COMPLETE

04/11/2025 12:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____