
Midmark M11 - v1.0.5
Total Cycles: 834

Name: RITTER2
Sterilizer ID: V2313228
02/24/2022 09:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.3 F 0.0
2:00 147.0 F 1.7
4:00 151.2 F 4.7
6:00 178.1 F 9.0
8:00 199.1 F 8.8
10:00 219.3 F 7.3
12:00 233.6 F 12.5
14:00 235.7 F 9.9
16:00 236.2 F 8.8
18:00 248.9 F 15.2
20:00 261.0 F 22.0
22:00 272.0 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.5 F 29.6
2:00 271.5 F 28.8
3:00 272.0 F 29.8
4:00 272.2 F 29.3

Min 271.2 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.1 F 0.0
5:00 202.1 F 0.0
10:00 204.5 F 0.0
15:00 203.0 F 0.0
20:00 196.8 F 0.0
25:00 204.0 F 0.0
30:00 220.0 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 22:32
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:00:45
CYCLE COMPLETE

02/24/2022 10:32 AM
ACCEPTED: X
REJECTED:
Operator: U02