
Midmark M11 - v1.0.5
Total Cycles: 2015

Name: RITTER2
Sterilizer ID: V2313228
02/06/2023 05:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.9 F 0.0
2:00 188.3 F 3.4
4:00 208.6 F 8.8
6:00 224.8 F 9.1
8:00 235.9 F 11.4
10:00 237.0 F 9.9
12:00 236.9 F 8.8
14:00 252.0 F 16.5
16:00 264.3 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.3 F 29.1
2:00 272.2 F 29.6
3:00 271.9 F 28.9
4:00 272.2 F 29.6

Min 271.4 F 28.7
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 227.5 F 0.0
10:00 226.6 F 0.0
15:00 223.3 F 0.0
20:00 230.5 F 0.0
25:00 220.4 F 0.0
30:00 221.3 F 0.0

Drying Complete

FILLING: 2:38
HEATING: 17:49
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:36
CYCLE COMPLETE

02/06/2023 06:47 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____