
Midmark M11 - v1.0.5
Total Cycles: 2763

Name: RITTER2
Sterilizer ID: V2313228
08/14/2023 04:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.7 F 0.0
2:00 172.8 F 0.0
4:00 200.7 F 0.6
6:00 214.8 F 5.1
8:00 228.6 F 9.7
10:00 233.7 F 9.3
12:00 241.7 F 12.4
14:00 241.2 F 11.4
16:00 240.7 F 10.9
18:00 253.9 F 17.6
20:00 265.3 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.1 F 28.9
2:00 272.3 F 29.4
3:00 271.3 F 28.5
4:00 272.2 F 29.1

Min 271.2 F 28.5
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.3 F 0.0
5:00 224.5 F 0.0
10:00 220.1 F 0.0
15:00 182.1 F 0.0
20:00 176.4 F 0.0
25:00 172.8 F 0.0
30:00 170.9 F 0.0
Drying Complete

FILLING: 1:59
HEATING: 21:46
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:01:20
CYCLE COMPLETE

08/14/2023 05:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____