
Midmark M11 - v1.0.5
Total Cycles: 727

Name: RITTER2
Sterilizer ID: V2313228
01/25/2022 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.1 F 0.0
2:00 200.7 F 4.4
4:00 210.4 F 10.5
6:00 223.2 F 6.9
8:00 238.0 F 13.4
10:00 241.1 F 12.2
12:00 241.9 F 11.8
14:00 255.6 F 19.0
16:00 267.3 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.1 F 29.2
2:00 272.5 F 29.7
3:00 271.4 F 29.0
4:00 272.3 F 29.5

Min 271.4 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 219.6 F 0.0
10:00 230.7 F 0.0
15:00 229.2 F 0.0
20:00 225.1 F 0.0
25:00 221.5 F 0.0
30:00 220.3 F 0.0
Drying Complete

FILLING: 2:52
HEATING: 17:17
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:18
CYCLE COMPLETE

01/25/2022 01:22 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____