
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001979
12:36 08/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:54
PRESSURIZING 2:58
STERILIZING 4:17
136.4 C 221kPa 4:17
136.4 C 225kPa 4:47
136.4 C 223kPa 5:17
136.2 C 222kPa 5:47
136.2 C 219kPa 6:17
136.4 C 223kPa 6:47
136.4 C 221kPa 7:17
136.4 C 224kPa 7:47
136.4 C 224kPa 7:47
VENTING 7:48
AIR DRYING 8:40
CYCLE COMPLETE 8:50

Digital Signature #
C9FDBFBB355B4CB1
