

-----  
Midmark M11 - v1.0.0  
Total Cycles: 601

Name: MIDMARK  
Sterilizer ID: V2178198  
01/13/2021 10:30 AM

BEGIN PACKS CYCLE  
Temp: 250 Degrees F  
Time: 30 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 127.1 F 0.0  
2:00 139.3 F 1.0  
4:00 152.5 F 3.2  
6:00 180.9 F 6.9  
8:00 203.0 F 2.4  
10:00 217.9 F 6.7  
12:00 223.2 F 4.6  
14:00 226.8 F 4.8  
16:00 237.0 F 9.1  
18:00 249.2 F 14.8

STERILIZING  
mm:ss Degrees PSI  
0:00 251.3 F 16.0  
1:00 252.2 F 16.2  
2:00 252.7 F 16.6  
3:00 251.3 F 16.0  
4:00 252.5 F 16.5  
5:00 251.2 F 16.1  
6:00 252.5 F 16.5  
7:00 251.2 F 16.0  
8:00 252.6 F 16.5  
9:00 251.7 F 16.1  
10:00 252.1 F 16.6  
11:00 252.3 F 16.4  
12:00 251.4 F 16.0  
13:00 252.3 F 16.6  
14:00 252.3 F 16.4  
15:00 251.6 F 16.1  
16:00 251.8 F 16.5  
17:00 252.4 F 16.5  
18:00 251.9 F 16.2  
19:00 251.2 F 16.1  
20:00 252.6 F 16.6  
21:00 252.2 F 16.4  
22:00 251.6 F 16.1  
23:00 251.5 F 16.4  
24:00 252.5 F 16.6  
25:00 252.1 F 16.3  
26:00 251.5 F 16.0  
27:00 251.8 F 16.5  
28:00 252.5 F 16.6  
29:00 252.1 F 16.3  
30:00 251.5 F 16.0

Min 251.1 F 15.9  
Max 252.8 F 16.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.0 F 0.0  
5:00 197.7 F 0.0  
10:00 200.3 F 0.0  
15:00 204.8 F 0.0  
20:00 200.6 F 0.0  
25:00 199.1 F 0.0  
30:00 194.5 F 0.0  
Drying Complete

FILLING: 0:59

HEATING: 18:58  
STERILIZING: 30:00  
VENTING: 2:53  
DRYING: 30:00  
TOTAL CYCLE: 01:22:50  
CYCLE COMPLETE

01/13/2021 12:34 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01