
Midmark M11 - v1.0.0
Total Cycles: 2127

Name: RITTER 1
Sterilizer ID: V2178198
07/14/2022 09:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.9 F 0.0
2:00 172.3 F 1.7
4:00 181.6 F 5.6
6:00 205.8 F 11.3
8:00 219.5 F 5.8
10:00 233.1 F 10.8
12:00 233.5 F 7.1
14:00 244.6 F 13.2
16:00 243.2 F 12.1
18:00 256.0 F 18.8
20:00 267.5 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.5
1:00 272.6 F 29.3
2:00 271.7 F 28.4
3:00 272.6 F 29.3
4:00 271.8 F 28.5

Min 271.4 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.1 F 0.0
5:00 216.7 F 0.0
10:00 227.6 F 0.0
15:00 227.2 F 0.0
20:00 225.2 F 0.0
25:00 224.4 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 21:12
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:56
CYCLE COMPLETE

07/14/2022 10:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____