
Midmark M11 - v1.0.5
Total Cycles: 3647

Name: RITTER2
Sterilizer ID: V2313228
05/27/2024 10:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.5 F 0.0
2:00 159.4 F 1.8
4:00 167.1 F 5.2
6:00 192.7 F 9.4
8:00 211.9 F 8.8
10:00 222.0 F 8.2
12:00 233.8 F 13.0
14:00 235.6 F 10.3
16:00 236.0 F 8.8
18:00 248.2 F 15.1
20:00 259.7 F 21.5
22:00 270.1 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.7 F 30.1
2:00 271.9 F 29.5
3:00 271.1 F 29.4
4:00 272.5 F 30.1

Min 271.0 F 29.2
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 209.3 F 0.0
10:00 198.7 F 0.0
15:00 195.2 F 0.0
20:00 193.4 F 0.0
25:00 210.0 F 0.0
30:00 217.2 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 22:58
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:01:50
CYCLE COMPLETE

05/27/2024 11:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____